





Collect the Autograft bone to locate the drill on the surface of operation area by pressing it slightly with saline irrigation. (Drilling Speed: 300rpm)

If use of ACM has accstomed, collect the bone in 80rpm condition without the irrigation.

Neobiotech Co., Ltd. 104, 24 / 102, 103, 104-1, 104-2, 105, 106, 205, 212, 312, 509, 510, 511, 10F, 36 Digital-ro 27-gil, Guro-gu, Seoul, Korea Tel 82-2-582-2885, Fax 82-2-582-2883 www.neobiotech.com / Made in Korea